

Receiving Report

Date: 17/03/15

Batch No: 137067

Supplier: Test Metric

Dart P/O: 3810

Packing Slip: Yes ☒ No ☐
 Invoice: Yes ☒ No ☐
 Receipt: Cash ☐ Cr ☒
 New Supplier Yes ☐ No ☒

Release Note Attached: Yes ☒ No ☐ N/A ☐
 Waybill Attached: Yes ☒ No ☐
 Shipment Complete: Yes ☒ No ☐ N/A ☐
 QC18 Inspection ☐ N/A ☒
 Work Order ☐ N/A ☒

Discrepancies

Part Number	Description	Quantity Ordered	Quantity Rec'd	Quantity Short	Quantity Inspected	Quantity Rejected	Comment / NCR Number

Initials of Receiver QC12 SD

Production/Admin:
 Date _____
 Received/Costing _____
 Initial _____

Location _____



Dart Aerospace Ltd.
1270 Aberdeen Street
Hawkesbury, ON K6A 1K7
Tel: 613 632 9577
Fax: 613 632 1053

PURCHASE ORDER

Purchase Order ID **PO35510**

Purchase Order Date 3/7/2017

PO Print Date 3/7/2017

Page Number 1 of 1

Order From :

VC-TM001

Ship To : DART AEROSPACE LTD

TESTMETRIC.

605 C CITE DES JEUNES

ST-LAZARE, QUEBEC J7T 2A7

1270 ABERDEEN

HAWKESBURY, ON K6A 1K7

CANADA

MAR 08 2017

Contact Name

Vendor Phone 450-424-7717

Buyer

Chantal Lavoie

Customer POID

Customer Tax # 10127-2607

Ship To Contact

Terms Net 30

Ship To Phone

Currency CAD

Ship Via:

FedEx Economy collect

FOB

FCA - (Free Carrier)

Ship Acct:

Line Nbr	Reference Vendor Part Number Line Comments Delivery Comments	Description/ Mfg ID	Req Date/ Taxable Promise Date	CD	Req Qty/ Unit of Measure	PO Unit Price	Extended Price
71350-20		PRE-SET TORQUE SCREWDRIVER CALIBRATION	3/17/2017		1.00	\$50.00	\$50.00
	PROCUREMENT QUALITY CLAUSES A044 CALIBRATION OF EQUIPMENT, CERTIFICATE OF CALIBRATION REQUIRED		Yes 3/17/2017		Each		

MAR 15 2017

Deliver To: ERIC.D

Line Total:

\$50.00

PO Total:

\$50.00

PO Instructions: PROCUREMENT QUALITY CLAUSES
A044 CALIBRATION OF EQUIPMENT, CERTIFICATE OF CALIBRATION REQUIRED

Note: Terms & Condition of Purchasing(Suppliers) and Procurement Quality Clauses are an integral part of our AS9100 requirements. To learn in detail, please visit www.dartaerospace.com for further explanation.

Change Nbr:

1

Change Date: 3/7/2017

**Instrument & Calibration Services**

605C Cite des Jeunes

St-Lazare, Qc, J7T 2A7

Phone: 450-424-7717 Fax: 450-424-4263

www.testmetric.com

PACKING SLIP

DATE: 13-Mar-17

Purchase Order Number
PO35510**Bill to:**

DART AEROSPACE

1270 Aberdeen Street

Hawkesbury, ON, K6A 1K7

Ship to:

DART AEROSPACE

1270 Aberdeen Street

Hawkesbury, ON, K6A 1K7

ORDER DATE	SHIP DATE	CONTACT	F.O.B.	ACCOUNT#	SHIP VIA
08-Mar-17	13-Mar-17	CHANTAL LAVOIE	YUL	DART	FEDEX

MODEL	DESCRIPTION	S/N	TYPE	ORDER QUANTITY	SHIP QUANTITY
21SP	Torque Screwdriver	0816402145	CAL	1	1
MAR 15 2017					

Signature: _____ Date: _____

Please contact the Customer Service department at 877-424-7717 with any questions or concerns.

Certificate No.: 886602



605C, Cité des Jeunes
St-Lazare, Qc, Canada J7T 2A7

Tél. 450-424-7717
Fax 450-424-4263
www.testmetric.com

CALIBRATION CERTIFICATE

Asset Number: 082016
MFG/Model Number: 21SP
Serial Number: 0816402145
Description: TORQUE SCREWDRIVER
Customer: DART AEROSPACE
1270 Aberdeen
Hawkesbury, ON K6A 1K7
Customer P.O.: PO35510
Remarks: IN SPEC'S NO ADJ MADE

MAR 15 2017

Set Torque C.W. in-oz	Minimum	Maximum	Reading 1	Reading 2	Reading 3
6.0	5.76	6.24	5.92	5.96	5.94
19.0	18.24	19.76	18.95	18.95	18.93
32.0	30.72	33.28	31.90	31.89	31.90
ACCURACY = $\pm 4\%$ C.W.					

Set Torque C.C.W. in-oz	Minimum	Maximum	Reading 1	Reading 2	Reading 3
ACCURACY = $\pm 6\%$ C.C.W.					

The Quality System of Testmetric is registered to the Quality System Standard ISO 9001-2008 Certificate Number CA09/55389. Testmetric is in compliance with ISO10012, ISO/IEC 17025 and ANSI/NCSL Z540-1-1994

At planned intervals measurement standards are calibrated by comparison to or measurement against national standards. Supporting documentation relative to traceability is on file and available for examination upon request.

This certificate cannot be reproduced except in full without written approval of Testmetric. This certificate pertains to only the asset listed above.

Laboratory Standards

Asset	Mfg/Model	Description	Cal Date	Cal Due Date
STD372	INT/102028	Torque Wrench Calibrator	31-Mar-16	31-Mar-17

Conditions of Calibration

Temperature/Humidity: 20 \pm 3°C / 30-50% RH
Cal Procedure: SCP-53
Cal Date: 13-Mar-17
Due Date: 13-Mar-18

Calibrated by: PB



Quality Assurance: